- (1) Reuse guidelines. Voluntary guidelines adopted by the AAMI ("Reuse of Hemodialyzers," second edition). Incorporation by reference of the AAMI's "Reuse of Hemodialyzers," second edition, 1993, was approved by the Director of the Federal Register in accordance with 5 U.S.C. 552(a) and 1 CFR part 51. If any changes in "Reuse of Hemodialyzers," second edition, are also to be incorporated by reference, a notice to that effect will be published in the FEDERAL REGISTER.
- (2) Procedure for chemical germicides. To prevent any risk of dialyzer membrane leaks due to the combined action of different chemical germicides, dialyzers are exposed to only one chemical germicide during the reprocessing procedure. If a dialyzer is exposed to a second germicide, the dialyzer must be discarded.
- (3) Surveillance of patient reactions. In order to detect bacteremia and to maintain patient safety when unexplained events occur, the facility—
- (i) Takes appropriate blood cultures at the time of a febrile response in a patient; and
- (ii) If pyrogenic reactions, bacteremia, or unexplained reactions associated with ineffective reprocessing are identified, terminates reuse of hemodialyzers in that setting and does not continue reuse until the entire reprocessing system has been evaluated.
- (b) Standard: Transducer filters. To control the spread of hepatitis, transducer filters are changed after each dialysis treatment and are not reused.
- (c) Standard: Bloodlines. If the ESRD facility reuses bloodlines, it must—
- (1) Limit the reuse of bloodlines to the same patient;
- (2) Not reuse bloodlines labeled for "single use only";
- (3) Reuse only bloodlines for which the manufacturer's protocol for reuse

- has been accepted by the Food and Drug Administration (FDA) pursuant to the premarket notification (section 510(k)) provision of the Food, Drug, and Cosmetic Act; and
- (4) Follow the FDA-accepted manufacturer's protocol for reuse of that bloodline.

[52 FR 36935, Oct. 2, 1987, as amended at 55 FR 18335, May 2, 1990; 60 FR 48044, Sept. 18, 1995]

§ 405.2160 Condition: Affiliation agreement or arrangement.

- (a) A renal dialysis facility and a renal dialysis center (see §405.2102(e)(2)) have in effect an affiliation agreement or arrangement with each other, in writing, for the provision of inpatient care and other hospital services.
- (b) The affiliation agreement or arrangement provides the basis for effective working relationships under which inpatient hospital care or other hospital services are available promptly to the dialysis facility's patients when needed. The dialysis facility has in its files documentation from the renal dialysis center to the effect that patients from the dialysis facility will be accepted and treated in emergencies. There are reasonable assurances that:
- (1) Transfer or referral of patients will be effected between the renal dialysis center and the dialysis facility whenever such transfer or referral is determined as medically appropriate by the attending physician, with timely acceptance and admission;
- (2) There will be interchange, within 1 working day, of the patient long-term program and patient care plan, and of medical and other information necessary or useful in the care and treatment of patients transferred or referred between the facilities, or in determining whether such patients can be adequately cared for otherwise than in either of such facilities; and
- (3) Security and accountability for patients' personal effects are assured.

§ 405.2161 Condition: Director of a renal dialysis facility or renal dialysis center.

Treatment is under the general supervision of a Director who is a physician. The physician-director need not

¹The publication entitled "Reuse of Hemodialyzers," second edition, 1993, is available for inspection at the HCFA Information Resources Center, 7500 Security Boulevard, Baltimore, MD 21244–1850 and the Office of the Federal Register, 800 North Capitol Street, NW., Suite 700, Washington, DC. Copies may be purchased from the Association for the Advancement of Medical Instrumentation, 3300 Washington Boulevard, Suite 400, Arlington, VA 22201–4598.